

# The intense Positron Source EPOS at ELBE Radiation Source of Research Center Rossendorf

**R. Krause-Rehberg<sup>1</sup>, G. Brauer<sup>2</sup>,**

<sup>1</sup>Martin-Luther-University Halle

<sup>2</sup>Research Center Rossendorf




TEKNILLINEN KORKEAKOULU  
TEKNISKA HÖGSKOLAN  
HELSINKI UNIVERSITY OF TECHNOLOGY



# The EPOS positron source at Research Center Rossendorf

- Main experiment in Rossendorf: Radiation source ELBE = **E**lectron **L**inac with high **B**rilliance and low **E**mittance
- Primary electron beam (40 MeV  $\times$  1 mA = 40 kW)
- Main goal: IR Free-electron Laser
- Very interesting time structure: cw-mode of short bunches




electron bunches

## EPOS = ELBE Positron Source

- Intense beam of slow (monoenergetic) positrons
- All relevant positron techniques for materials research (positron lifetime, Coincidence Doppler broadening, AMOC)
- EPOS is external facility of Martin-Luther-University Halle at Research center Rossendorf
- User-dedicated facility
- Remote controlled via internet
- Financing by University Halle, Land Sachsen-Anhalt and European Community



# Ground map of the ELBE hall




# Positron Lab

- positron lab in ELBE hall already available

X-ray Lab

Positron Lab

concrete screening of Cave 111b  
(location of  $e^+$  converter)







3,2 m concrete screening  
of Cave 111b

cable tunnel to be used for  $e^+$   
beamline

photo taken in November 2003



# Second timing mode needed for long lifetimes (porous material)





Result of Fit:

Repetition time	Lifetime	Intensity
616 ns	141 ns	5,1 %
308 ns	124 ns	4,9 %
154 ns	75 ns	3,3 %
77 ns	35 ns	1,7 %



# Converter





MCNP-Simulationen A. Rogov und K. Noack (FZR)





# Directly water-cooled Electron-Positron Converter

- first attempt: porous W (30 % porosity) -> too low water flux at 10 bar
- stack of 50 pieces W-foils 0,1 mm separated by 0,1 mm -> 13,5 l water at 1,5 bar
- foils cut by IR-laser in our workshop




# Converter Chamber





**Beam Dump**  
≈ 250 kG  
50 cm diameter  
50 cm long  
pure Al (99.7%)


# Simulation of Positron distribution



# Simulation of Energy deposition

**Al beam dump 21 kW**  
(made of 5N-purity)

**W target**  
**14 kW**




EPOS Density Energy Deposition (in MeV/cm<sup>3</sup>) for Distance = 10cm

**↑ primary beam**





# Simulation of Positron Energy Distribution

primary electron beam 40 MeV



# Simulation of expected $\gamma$ and n dose




Screening by lead blocks, Polyethylene bricks and heavy concrete





# Moderator



Positronenfluss durch den Moderator (d=20mm) für unterschiedliche Positronen-Energiebereiche und unterschiedliche Winkel (normiert auf 0°)

Intensität bei 0 °:

E < 250 keV:  $6,9 \times 10^9$


E > 250 keV:  $8,4 \times 10^{11}$

MCNP-Simulationen A. Rogov und K. Noack (FZR)




# Simulation of positron extraction

- simulation done by EGUN
- area of 20 mm diameter at moderator is used and squeezed to about 2 mm



# EPOS scheme



# Magnetic Beam Guidance


Magnetic field of 75 Gauss


45 coils but only  
5 different currents  
5 Power supplies  
2.7 kW Power together

→ maximum change 6 G  
→ Gradient < 0,11 G/mm

Steering coils

30 coils with different  
(computer-driven) currents








- Simulation of bunching by POSBUNCH
- C++ author: V. Bondarenko
- source code available on request

# Simulation of Buncher Voltages

Both buncher RF-voltage amplitudes and the drift path energy must be adjusted for each beam energy for optimum time resolution




# Chopper






# Detector system

- **3 experiments:** lifetime spectroscopy (8 BaF<sub>2</sub> detectors); Doppler coincidence (2 Ge detectors), and AMOC (1 Ge and 1 BaF<sub>2</sub> detector)
- **digital detection system:**
  - lifetime: almost nothing to adjust; time scale exactly the same for all detectors; easy realization of coincidence
  - Doppler: better energy resolution and pile-up rejection expected
  - pulse-shape discrimination improves spectra quality



# Lifetime detector system



- Linux PC's
- dual XEON 64bit/3GHz
- two-processor system
- no monitors, no harddiscs
- 4 GB memory each



# Time Schedule

	1. Year	2. Year	3. Year
Laboratory	██████████		
Simulation e <sup>+</sup> converter	██████████		
Simulation beam	████████████████		
Converter chamber and vacuum system in tunnel	██████████████		
Screening of converter chamber		██████	
First chopper / buncher		██████	
Test converter / beam transport		██	
Vacuum system completion		██████	
Conventional source chamber		██████████████	
2. Chopper / buncher		██████████	
Sample chamber			██████████
Completion of beam electronics			██████
Test transport system			██████████
Detector system and software	██		
Automation			██
Software lifetime / Doppler spectra			████████████████████
Optimization of time resolution			████████████████████



# EPOS - Applications

## Variety of applications in all field of materials science:

- defect-depth profiles due to surface modifications and ion implantation
- tribology (mechanical damage of surfaces)
- polymer physics (pores; interdiffusion; ...)
- low-k materials (thin high porous layers)
- defects in semiconductors, ceramics and metals
- epitaxial layers (growth defects, misfit defects at interface, ...)
- fast kinetics (e.g. precipitation processes in Al alloys; defect annealing; diffusion; ...)
- radiation resistance (e.g. space materials)
- many more ...

